

# ELEKTROTECHNICKÝ ZKUŠEBNÍ ÚSTAV



ELECTROTECHNICAL TESTING INSTITUTE - CZECH REPUBLIC  
ELEKTROTECHNISCHE PRÜFANSTALT - TSCHJECHISCHE REPUBLIK  
INSTITUT ELECTROTECHNIQUE D'ESSAIS - RÉPUBLIQUE TCHÈQUE  
ЭЛЕКТРОТЕХНИЧЕСКИЙ ИСПЫТАТЕЛЬНЫЙ ИНСТИТУТ - ЧЕШСКАЯ РЕСПУБЛИКА

Pod Lisem 129, 171 02 Praha 8 - Troja

## CERTIFICATE

No.: 1170681

**Product:** Active flexible sensor for measurement of alternating current

**Type:** AMOS M long

**Rating:** 230 V; 30, 100, 300, 1000, 3000 A, Safety class II, CAT IV/300 V  
Supply voltage  $\pm 3$  V to  $\pm 12$  V, Output voltage 1,0 V AC for Inom

**Ordering firm:** MEgA - Měřicí Energetické Aparáty, a.s.  
Česká 390, 664 31 Česká, Czech Republic

**Manufacturer:** MEgA - Měřicí Energetické Aparáty, a.s.  
Česká 390, 664 31 Česká, Czech Republic

**Factories:** MEgA - Měřicí Energetické Aparáty, a.s.  
Česká 390, 664 31 Česká, Czech Republic

**Trade mark:**

**The test results are stated in the test-report No.:** 701806-01/01 of: 15.09.2017, 701806-01/02 of: 27.07.2017

**A sample of the product was found to be in conformity with:**

EN 61010-1:10  
EN 61010-2-032:12  
EN 61326-1:13

**Other data:**

Certificate was issued on the basis of fulfillment of requirements of the "EZÚ certificate" certification scheme and on the basis of agreement No. 701806 between the client and the Electrotechnical Testing Institute.

**Compliance of the product with mentioned standards and regulations ensures compliance of the product with essential requirements of Government Order No. as amended and the certificate may be used as a supporting document for the EU Declaration of Conformity under Act No. 90/2016 Coll., on Conformity Assessment of Products When Made Available on the Market, as amended.**

**The validity of the certificate is limited to:** 30.09.2020

20.09.2017

Prague

  
Mgr. Miroslav Sedláček  
Head of Certification Body



Stamp



701806-01