

ELEKTROTECHNICKÝ ZKUŠEBNÍ ÚSTAV



ELECTROTECHNICAL TESTING INSTITUTE - CZECH REPUBLIC
ELEKTROTECHNISCHE PRÜFANSTALT - TSCHECHISCHE REPUBLIK
INSTITUT ELECTROTECHNIQUE D'ESSAIS - RÉPUBLIQUE TCHÉQUE
ЭЛЕКТРОТЕХНИЧЕСКИЙ ИСПЫТАТЕЛЬНЫЙ ИНСТИТУТ - ЧЕШСКАЯ РЕСПУБЛИКА

Pod lisem 129/2, 171 02 Praha 8 - Troja

CERTIFICATE

No.: 1210305

Product: Smart split core current transformer

Type: SMART PTD

Rating: Current: 200; 400; 600 nebo 900 A / 20 mA; CAT IV/300 V AC, Safety Class II

Ordering firm: MEgA - Měřicí Energetické Aparáty, a.s.
Česká 390, 664 31 Česká, Czech Republic

Manufacturer: MEgA - Měřicí Energetické Aparáty, a.s.
Česká 390, 664 31 Česká, Czech Republic

Factories: MEgA - Měřicí Energetické Aparáty, a.s.
Česká 390, 664 31 Česká, Czech Republic

Trade mark:

The test results are stated in the test-report No.: 210986-01/01 of: 26.05.2021, 210986-01/02 of: 26.05.2021

A sample of the product was found to be in conformity with:
EN 61010-1:10+A1:16, EN 61010-2-032:12

Other data:

The validity of the certificate is limited to: 30.05.2024

31.05.2021

Prague

Mgr. Miroslav Sedláček
Head of Certification Body



Stamp



210986-01